

File No. 16/06/2015-NISE
National Institute of Solar Energy
Gwalpahari Gurgaon
(An Autonomous institute of ministry of new & Renewable Energy)

Date: 03.12.2015

31st Dec 2015

CORRIGENDUM

Sub: Regarding procurement of power supply for indoor testing (PVTF LAB, ETL LAB, BTL LAB).

In the section "B", Specifications of DC Power supply (at page no. 4), it was written as:

1. Sr. No -4, Meter accuracy for current: 0.01 % for the power rating (Types) 1.2 KW and 600 W .
2. Sr.no -10, Environmental for operating temperature: 0-50°C for the power rating (Types) 1.2 KW, 600 W & 2 KW.

May be read as:

1. Sr. No -4, Meter accuracy for current: 0.1 % for the power rating (Types) 1.2 KW and 600 W
2. Sr.no -10, Environmental for operating temperature: 0-40°C for the power rating (Types) 1.2 KW, 600 W & 2 KW.

All other terms and conditions are remaining same.

The last date for submission of tender bid is extended up to 30 December, 2015.

Rajesh Kumar
8/12/15

Dr. Rajesh Kumar
(DDG, NISE)

Copy To:

1. PA TO DG
2. File No. 16/06/2015-NISE